



## AMP One Step Amphetamine Test Device (Urine) Package Insert

Cat: AMP-102  
Version: Z

Format: Device  
Effective Date: 2020-07

For professional *in vitro* diagnostic use only.

### INTENDED USE

The AMP One Step Amphetamine Test Device (Urine) is a lateral flow chromatographic immunoassay for the detection of Amphetamine in urine at a cut-off concentration of 1000 ng/mL. This test will detect other related compounds, please refer to the Analytical Specificity table in this package insert

This assay provides only a qualitative, preliminary analytical test result. A more specific alternate chemical method must be used in order to obtain a confirmed analytical result. Gas chromatography/mass spectrometry (GC/MS) is the preferred confirmatory method. Clinical consideration and professional judgment should be applied to any drug of abuse test result, particularly when preliminary positive results are used.

### INTRODUCTION

Amphetamine is a Schedule II controlled substance available by prescription (Dexedrine®) and is also available on the illicit market. Amphetamines are a class of potent sympathomimetic agents with therapeutic applications. They are chemically related to the human body's natural catecholamines: epinephrine and norepinephrine. Acute higher doses lead to enhanced stimulation of the central nervous system and induce euphoria, alertness, reduced appetite, and a sense of increased energy and power. Cardiovascular responses to Amphetamines include increased blood pressure and cardiac arrhythmias. More acute responses produce anxiety, paranoia, hallucinations, and psychotic behavior. The effects of Amphetamines generally last 2-4 hours following use, and the drug has a half-life of 4-24 hours in the body. About 30% of Amphetamines are excreted in the urine in unchanged form, with the remainder as hydroxylated and deaminated derivatives.

The AMP One Step Amphetamine Test Device (Urine) is a rapid urine screening test that can be performed without the use of an instrument. The test utilizes a monoclonal antibody to selectively detect elevated levels of Amphetamines in urine. The AMP One Step Amphetamine Test Device (Urine) yields a positive result when Amphetamines in urine exceed 1,000 ng/mL.

### PRINCIPLE

The AMP One Step Amphetamine Test Device (Urine) is a rapid chromatographic immunoassay based on the principle of competitive binding. Drugs which may be present in the urine specimen compete against the drug conjugate for binding sites on the antibody.

During testing, a urine specimen migrates upward by capillary action. Amphetamine, if present in the urine specimen below 1,000 ng/mL, will not saturate the binding sites of the antibody coated particles in the Test Strip. The antibody coated particles will then be captured by immobilized Amphetamine conjugate and a visible colored line will show up in the test line region. The colored line will not form in the test line region if the Amphetamine level exceeds 1,000 ng/mL because it will saturate all the binding sites of anti-Amphetamine antibodies.

A drug-positive urine specimen will not generate a colored line in the test line region because of drug competition, while a drug-negative urine specimen or a specimen containing a drug concentration less than the cut-off will generate a line in the test line region. To serve as a procedural control, a colored line will always appear at the control line region indicating that proper volume of specimen has been added and membrane wicking has occurred.

### REAGENTS

The test contains drug-bovine protein antigen conjugate on the membrane and the conjugate pad of each test contains monoclonal anti-drug antibody.

### KIT COMPONENTS

Individually packed test Devices	Each Device contains a strip with colored conjugates and reactive reagents pre-spread at the corresponding regions. For operation instruction.
----------------------------------	---

### Package insert

### MATERIALS REQUIRED BUT NOT PROVIDED

Specimen collection container	For specimens collection use.
Timer	For timing use.

### PRECAUTIONS

- For professional *in vitro* diagnostic use only.
- Do not use after expiration date indicated on the package. Do not use the test if its foil pouch is damaged. Do not reuse tests.
- This kit contains products of animal origin. Certified knowledge of the origin and/or sanitary state of the animals does not totally guarantee the absence of transmissible pathogenic agents. It is therefore, recommended that these products be treated as potentially infectious, and handled observing the usual safety precautions (do not ingest or inhale).
- Avoid cross-contamination of specimens by using a new specimen collection container for each specimen obtained.
- Read the entire procedure carefully prior to performing any tests.
- Do not eat, drink or smoke in the area where the specimens and kits are handled. Handle all specimens as if they contain infectious agents. Observe established precautions against microbiological hazards throughout the procedure and follow the standard procedures for proper disposal of specimens. Wear protective clothing such as laboratory coats, disposable gloves and eye protection when specimens are assayed.
- Humidity and temperature can adversely affect results.
- The used testing materials should be discarded in accordance with local, state and/or federal regulations.

### STORAGE AND STABILITY

- The kit should be stored at 2-30°C until the expiry date printed on the sealed pouch.
- The test must remain in the sealed pouch until use.
- Do not freeze.**
- Cares should be taken to protect components in this kit from contamination. Do not use if there is evidence of microbial contamination or precipitation. Biological contamination of dispensing equipments, containers or reagents can lead to false results.

### SPECIMEN COLLECTION AND STORAGE

- The urine specimen must be collected in a clean and dry container. Urine collected at any time of the day may be used. Urine specimens exhibiting visible particles should be centrifuged, filtered, or allowed to settle to obtain clear specimen for testing.
- Collected urine specimens must be put in clear and dry containers. Ensure that a sufficient quantity of the specimen is collected to allow submerging the dipping area of the strip.
- Perform the testing immediately after the specimen collection. Do not leave the specimens at room temperature for prolonged periods. Specimens may be stored at 2-8°C for up to 48 hours. For long term storage, specimens should be kept below -20°C.
- Bring specimens to room temperature prior to testing. Frozen specimens must be completely thawed and mixed well prior to

testing. Avoid repeated freezing and thawing of specimens.

- Pack the specimens in compliance with applicable regulations for transportation of etiological agents, in case they need to be shipped.

### PROCEDURE

**Bring tests, specimens and/or controls to room temperature (15-30°C) before use.**

- Bring the pouch to room temperature before opening it. Remove the test device from the sealed pouch and use it as soon as possible.
- Place the test device on a clean and level surface. Hold the dropper vertically and **transfer 3 full drops of urine** (approx. 100 µL) to the specimen well (S) of the test device, and then start the timer. Avoid trapping air bubbles in the specimen well (S).
- Wait for the colored line(s) to appear. **Read results at 5 minutes.** Do not interpret the result after 10 minutes.

### INTERPRETATION OF RESULTS

#### POSITIVE RESULT:

C  
T

Only one colored band appears in the control region (C). No apparent colored band appears in the test region (T).

#### NEGATIVE RESULT:

C  
T

Two colored bands appear on the membrane. One band appears in the control region (C) and another band appears in the test region (T).

#### INVALID RESULT:

C  
T

Control band fails to appear. Results from any test which has not produced a control band at the specified reading time must be disregarded. Please review the procedure and repeat with a new test. If the problem persists, discontinue using the kit immediately and contact your local distributor.

#### NOTE:

- The intensity of the color in test region (T) may vary depending on the concentration of aimed substances present in the specimen. Therefore, any shade of color in the test region should be considered negative. Besides, the concentration level can not be determined by this qualitative test.
- Insufficient specimen volume, incorrect operation procedure, or performing expired tests are the most likely reasons for control band failure.

### QUALITY CONTROL

- Internal procedural controls are included in the test. A colored band appearing in the control region (C) is considered an internal positive procedural control. It confirms sufficient specimen volume and correct procedural technique.
- External controls are not supplied with this kit. It is recommended that positive and negative controls be tested as a good laboratory practice to confirm the test procedure and to verify proper test performance.

### LIMITATIONS OF THE TEST

- The AMP One Step Amphetamine Test Device (Urine) provides only a qualitative, preliminary analytical result. A secondary analytical method must be used to obtain a confirmed result. Gas chromatography/mass spectrometry (GC/MS) is the preferred confirmatory method.<sup>1,2</sup>
- It is possible that technical or procedural errors, as well as other interfering substances in the urine specimen may cause erroneous results.
- Adulterants, such as bleach and/or alum, in urine specimens may produce erroneous results regardless of the analytical method used. If adulteration is suspected, the test should be repeated with another urine specimen.
- A positive result indicates presence of the drug or its metabolites but does not indicate level of intoxication, administration route or

concentration in urine.

5. A negative result may not necessarily indicate drug-free urine. Negative results can be obtained when drug is present but below the cut-off level of the test.
6. Test does not distinguish between drugs of abuse and certain medications.

**PERFORMANCE CHARACTERISTICS**

**A. Accuracy**

118 clinical urine specimens were analyzed by GC-MS and by the AMP One Step Amphetamine Test Device (Urine). Each test was performed by three operators. Samples were divided by concentration into five categories: negative, less than half the cutoff, near cutoff negative, near cutoff positive, and high positive. Results were as follows:

Method		GC/MS					
The AMP One Step Amphetamine Test Device		Ne g.	Neg. (< -50% cutoff )	Near cutoff neg. (-50% cutoff to cutoff)	Near cutoff pos. (cutoff to +50% cutoff)	Pos. (> +50 % cutoff )	% agree ment with GC/MS
AMP 1000	Positive	0	0	1	15	22	97.37 %
	Negative	51	14	14	1	0	98.75 %

**B. Precision**

A study was conducted at three physician offices for Amphetamine (1000 ng/mL) by professional operators using three different lots of product to demonstrate the within run, between run and between operator precision. An identical panel of coded specimens, containing drugs at the concentration of ± 50% and ± 25% cut-off level, was labeled as a blind and tested at each site. The results are given below:

Drug Conc.	n per site	Site A		Site B		Site C	
		N	P	N	P	N	P
Negative	10	10	0	10	0	10	0
-50% Cut-off	10	10	0	10	0	10	0
-25% Cut-off	10	9	1	9	1	10	0
+25% Cut-off	10	1	9	0	10	1	9
+50% Cut-off	10	0	10	0	10	0	10

**C. Effect of Urinary Specific Gravity**

Fifteen (15) urine samples of normal, high, and low specific gravity ranges (1.000-1.037) were spiked with drugs at 50% below and 50% above cut-off levels respectively. The AMP One Step Amphetamine Test Device (Urine) was tested in duplicate using fifteen drug free urine and spiked urine samples. The results demonstrate that varying ranges of urinary specific gravity do not affect the test results.

**D. Effect of Urinary pH**

The pH of an aliquoted negative urine pool was adjusted to a pH range of 5 to 9 in 1 pH unit increments and spiked with drugs at 50% below and 50% above cut-off levels. The spiked, pH adjusted urine was tested with the AMP One Step Amphetamine Test Device (Urine). The results demonstrate that varying ranges of pH do not interfere with the performance of the test.

**E. Cross-Reactivity**

The following tables list the concentrations of compounds (ng/mL) above which the AMP One Step Amphetamine Test Device (Urine) identified positive results at 5 minutes.

	(ng/mL)
l-Amphetamine	50000
d,l-Amphetamine	3000
d-Amphetamine	1000
(±)3,4-Methylenedioxy amphetamine(MDA)	2000
Phentermine	3000

**F. Non Cross-Reacting Compounds**

The following compounds yielded negative results up to a concentration of 100 µg/mL:

4-Acetamidophenol	Gatifloxacin	Penfluridol
Acetaminophen	Gemfibrozil	Penicillin G potassium salt
Acetylsalicylic Acid	Gentisic Acid	Penicillin G sodium salt
Albumin	Glclazide	Perphenazine
Amoxicillin	Glipizide	Phenacetin
Ampicillin	Glyburide	Phenelzine Sulfate
Ampicillin trihydrate	Guaiaacol	Phenothiazine
Aspartame	Guaifenesin	2-Phenylethylamine
Atropine	Hemoglobin	Pioglitazone
Baclofen	Hydralazine HCl	Piracetam
Benzoic Acid	Hydrochlorothiazide	Pravastatin sodium
Berberine Chloride	Hydrocortisone	Prednisone
Hydrate	Ibuprofen	Procaine
Bilirubin	Isoprenaline	Promethazine
Caffeine	Ketoconazole	hydrochlorine
Cephalexin	Ketoprofen	6-Propyl-2-thiouracil
Cephhradine	Lamotrigine	Pyridoxine
Chloral hydrate	L-Ascorbic acid	Pyrilamine Maleate
Chloramphenicol	Levofloxacin	Pyrogallac
Chlorpheniramine	Lidocaine	Quetiapine Fumarate
Maleate	Lidocaine Monohydrate	Quinine
Chlorpromazine	Lisinopril Dihydrate	Quinolinic acid
Cholesterol	Lithium carbonate	R,R(-)-Pseudoephedrine
Ciprofloxacin hydrate	Loperamide	Ranitidine base
Clarithromycin	Loratadine	Ranitidine
Clonidine solution	L-Thyroxine sodium	Riboflavin
Creatinine	Maprotiline	Rifampicin
D(-)-Norgestrel	Meprobamate	Risperidone
d,l-Propranolol	Minocycline	Salicylic acid
Deoxycorticosterone	Mosapride Citrate	Sertraline HCl
Dextromethorphan solution	Nalidixic acid	Simvastatin
Diclofenac	Naloxone HCl	Sodium 2-Propylvalerate
Diffunisal	Naltrexone HCl	Sulfamethazine
Digoxin	Naproxen	Sulindac
4-Dimethyl-aminoantipyrine	Nicotinamide	Tetracycline
Diphenhydramine	Nicotinic acid	Tetrahydrozoline
5,5-Diphenylhydantoin	Nifedipine	Theophylline
D-Lactose monohydrate	Nimodipine	Thiamine
D-Leucyl-L-tyrosine	Norethisterone Acetate	Thioridazine solution
Hydrate	Norfloracin Nicotinic	Tolbutamide
Dopamine	Noscapine	Topiramate
Droperidol	(±)-Octopamine	2,4,7-Triamino-6-Phenylpteridine
Enalapril Maleate	Ofloxacin	Trimethoprim
Erythromycin	Olanzapine	Tryptamine
Estradiol	Oxalic acid, anhydrous	Tyramine
Estrone		Uric acid
Ethyl 4-aminobenzoate		
Fluoxetine		

Fotemustine	Oxolinic acid	(±)-Verapamil
Furosemide	Paliperidone	Vitamin B1
Gabapentin	Pantoprazole sodium	Zomepirac

**LITERATURE REFERENCES**

- Baselt RC. Disposition of Toxic Drugs and Chemicals in Man. 2nd Ed. Biomedical Publ., Davis, CA. 1982; 488
- Hawks RL, CN Chiang. Urine Testing for Drugs of Abuse. National Institute for Drug Abuse (NIDA), Research Monograph 73, 1986

**Index of Symbols**

	Consult Instruction for use		Tests per kit		Do not use if package is damaged
	For in vitro diagnostic use only		Use by date		Do not reuse
	Store between 2-30°C		Lot Number		Catalogue number
	Keep away from sunlight		Keep dry		Manufacturer
	Caution		Date of manufacture		Authorized Representative

**Hangzhou Sejoy Electronics& Instruments Co.,Ltd.**  
Area C, Building 2, No.365, Wuzhou Road,Yuhang Economic Development Zone, Hangzhou City 311100 Zhejiang China  
Website: [www.sejoy.com](http://www.sejoy.com)

**Shanghai International Holding Corp. GmbH (Europe)**  
Eiffestrasse 80, 20537 Hamburg, Germany

